

## Diabetic Foot Ulcer (DFU) application



PICO SNPWT may be used on wounds up to 4.5cm deep. For wounds 0.5cm–4.5cm deep, a filler may be required. The following fillers may be used: ACTICOAT<sup>o</sup> FLEX, foam and/or gauze.\* Prepare the wound per local protocol. NO-STING-SKIN-PREP may be used to help the dressing maintain a seal.



Remove the backing from the dressing and apply to the foot. Keep the port away from the wound site and high enough from the heel that it will not rub on footwear. Position the port uppermost from the wound.



RENASYS<sup>o</sup> GEL patches may be used in difficult anatomical areas to help achieve or maintain a seal.



Connect the pump to the dressing by twisting the two halves of the connector together, insert the batteries and press the orange button to start therapy.



Assist the dressing to get good contact with the wound/skin by following the dressing with your hand while the pump is evacuating air from the dressing. Once the negative pressure has been successfully delivered, apply the fixation strips to the border of the dressing. The strips may be cut to make them easier to apply.

### Advanced Wound Management Customer Care Center

Smith+Nephew  
Fort Worth, TX 76109  
USA

T 800 876-1261  
F 727 392-6914

©2022 Smith+Nephew  
\*Trademark of Smith+Nephew  
737-5174-20221017



[www.possiblewithpico.com](http://www.possiblewithpico.com)

[www.smith-nephew.com](http://www.smith-nephew.com)

[Watch Application Video](#)

Smith+Nephew does not provide medical advice. It is the responsibility of clinicians to determine and utilize products and techniques appropriate for each of their individual patients. This is not the full application instructions, for more details please refer to the Instructions for Use (IFU). For detailed product information, including indications for use, contraindications, precautions and warnings, including the magnet warning, please consult the product's applicable IFU prior to use.

\*ACTICOAT FLEX 3 and ACTICOAT FLEX 7 are only approved for use with NPWT for up to 3 days.

**PRECAUTION:** The use of negative pressure presents a risk of tissue ingrowth into foam when this is used as a wound filler. When using a foam filler with PICO, tissue ingrowth may be reduced by using a non-adherent wound contact layer or by increasing the frequency of dressing changes.